



UNIVERSITY

MEMORIAL UNIVERSITY OF NEWFOUNDLAND

SCHOOL OF PHARMACY

OPEN CALL FOR BIDS FOR

1 (ONE) MICRO CT SCANNER

TPH-044-24

ADDENDUM # 1

Date: October 11, 2024

Purpose of Addendum # 1:

In Appendix A under Mandatory Specifications (Equipment) item # 2 currently reads:

“The system must be capable of a spatial 3D resolution of at least **5 um**, confirmed with independent resolution test, such as micro-CT bar Pattern NANO phantom.”

It should read:

“The system must be capable of a spatial 3D resolution of at least **7 um**, confirmed with independent resolution test, such as micro-CT bar Pattern NANO phantom.”

In Appendix A under Background Information, Description, item # 1 currently reads:

“The system should have capacity for imaging below **5 um**.”

It should read;

“The system should have capacity for imaging below **7 um**.”